VECTOR LABORATORIES

part of Maravai LifeSciences

CERTIFICATE OF ANALYSIS

Product Name	**CON A, Agarose**
Product Catalog No.	**AL-1003-10**
Unit Size	**10 ml**
Lot No.	**ZJ0429**
Lot Expiration Date	**April 30, 2024**

ASSAY	SPECIFICATION	LOT DATA
DYNAMIC BINDING CAPACITY	≥ 3.0 mg of test protein per mL of settled beads	PASS

Quality Assurance

05/09/22

Date: 05/03/2022

Vector Laboratories on behalf of Maravai LifeSciences, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.

Vector Laboratories, Inc.
30 Ingold Road
Burlingame, CA 94010 USA
Phone: (650) 697-3600
vectorlabs.com