

# Certificate of Analysis



**Product Name** PHA-L, Biotinylated  
**Product Catalog No.** B-1115-2  
**Unit Size** 2 mg  
**Lot No.** ZK0222  
**Lot Expiration Date** February 28, 2025

ASSAY	SPECIFICATION	LOT DATA
ACTIVITY	≥ 85%	PASS
BIOTINYLATION	≥ 97%	PASS
TISSUE TEST	EQUIVALENT TO CONTROL	PASS

Quality Assurance

Date: 02/23/2023

*Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.*