

Certificate of Analysis



Product Name ChromaLINK® One-Shot™ Antibody Biotinylation Kit
Product Catalog No. B-9007-009K
Unit Size 1 kit
Kit Lot No. ZK1019
Kit Expiration Date January 31, 2025

ASSAY	SPECIFICATION	LOT DATA
FUNCTIONAL TEST	Incorporates 3-8 biotin molecules per antibody molecule when kit is used as directed to label a single 100 ug quantity of antibody at 1 mg/ml.	PASS

COMPONENTS	LOT #
ChromaLINK® Biotin	ZK0907
Anhydrous DMF	ZK0905
7MWCO Zeba™ Columns	ZK0906
1X Modification Buffer	ZK0907
1X PBS	ZK0901
1M Tris HCl	ZK0906
Collection Tubes	ZK0908
Biotinylated Bovine IgG Control	ZK1012

Note: The MSR for the Biotinylated Bovine IgG Control can now be found in the B-9007-009K User Guide.

Quality Assurance *Hayden Kane*

Date: *10/23/2023*

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.