

Certificate of Analysis



Product Name ChromaLINK® One-Shot™ Antibody Biotinylation Kit
Product Catalog No. B-9007-009K
Unit Size 1 kit
Kit Lot No. ZL0823
Kit Expiration Date March 31, 2025

ASSAY	SPECIFICATION	LOT DATA
FUNCTIONAL TEST	Incorporates 3-8 biotin molecules per antibody molecule when kit is used as directed to label a single 100 ug quantity of antibody at 1 mg/ml.	PASS

COMPONENTS	LOT #
ChromaLINK® Biotin	ZL0823
Anhydrous DMF	ZL0730
7MWCO Zeba™ Columns	ZL0318
1X Modification Buffer	ZL0731
1X PBS	ZL0730
1M Tris HCl	ZL0731
Collection Tubes	ZL0614
Biotinylated Bovine IgG Control	ZL0816

Note: The MSR for the Biotinylated Bovine IgG Control can now be found in the B-9007-009K User Guide.

Quality Assurance Hayrett Kaur

Date: 09/19/2024

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.