

Certificate of Analysis



Product Name ChromaLINK® Biotin Protein Labeling Kit
Product Catalog No. B-9007-105K
Unit Size 1 kit
Kit Lot No. ZJ1020
Kit Expiration Date March 31, 2024

ASSAY	SPECIFICATION	LOT DATA
FUNCTIONAL TEST	Incorporates 3-8 biotin molecules per antibody when kit is used as directed to label Bovine IgG at 1 mg/ml	PASS

COMPONENTS	LOT #
ChromaLINK® Biotin	ZJ1017
10X Modification Buffer	ZJ0826
Anhydrous DMF	ZJ1013
0.5 mL Zeba™ Columns	ZJ1019
10X PBS	ZJ0825
2 mL Collection Tubes	ZJ1019
1.5 mL Storage Tubes	ZJ1019
Bovine IgG Control	ZJ1012
Biotinylated Bovine IgG Control	ZJ1013

Note: The MSR for the Biotinylated Bovine IgG Control can now be found in the B-9007-105K User Guide.

Quality Assurance

Date: 10/24/2022

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.