

Certificate of Analysis



Product Name ChromaLINK® Digoxigenin One-Shot™ Antibody Labeling Kit
Product Catalog No. B-9014-009K
Unit Size 1 kit
Kit Lot No. ZL0703
Kit Expiration Date April 30, 2025

ASSAY	SPECIFICATION	LOT DATA
FUNCTIONAL TEST	Incorporates 2-8 digoxigenin molecules per antibody when kit is used as directed to label 100 ug of antibody	PASS

COMPONENTS	LOT #
0.5 mL Zeba™ Columns	ZL0611
Sulfo ChromaLINK® Digoxigenin	ZL0319
Anhydrous DMF	ZL0416
1X Modification Buffer	ZL0306
1X PBS	ZL0614
1M Tris HCl	ZL0315
2 mL Collection Tubes	ZL0614
DIG-Labeled Bovine IgG Control	ZL0315

Note: The MSR for the DIG-Labeled Bovine IgG Control can now be found in the B-9014-009K User Guide.

Quality Assurance Hayrett Kaur

Date: 07/09/2024

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.