

Certificate of Analysis



Product Name High Capacity Streptavidin Magnetic Beads
Product Catalog No. CCT-1497-1
Unit Size 1 mL
Lot No. ZL0816
Retest Date August 31, 2026

Test Method	Lot Data	Specification	Result
Binding Capacity	12.62 nmol Biotin/mg	≥ 12 nmol Biotin/mg	PASS

Quality Assurance

M. Almdalgaard

Date

8/21/24

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis