

Certificate of Analysis



Product Name High Capacity Streptavidin Magnetic Beads
Product Catalog No. CCT-1497-5
Unit Size 5 mL
Lot No. ZL0702
Retest Date July 31, 2026

Test Method	Lot Data	Specification	Result
Binding Capacity	16.3 nmol Biotin/mg	≥ 12 nmol Biotin/mg	PASS

Quality Assurance

Hayrett Kaur

Date

08/26/2024

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis