

# Certificate of Analysis



**Product Name** Glycoprotein Eluting Solution (Mannose/Glucose)  
**Product Catalog No.** ES-1100-100  
**Unit Size** 100 ml  
**Lot No.** ZL0530  
**Lot Expiration Date** May 31, 2026

ASSAY	SPECIFICATION	LOT DATA
TLC	Rf 0.45-0.67	PASS
REFRACTIVE INDEX	1.1266 - 1.6900	PASS

Quality Assurance Harpreet Kaur

Date: 05/31/2024

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.