

Certificate of Analysis



Product Name UEA I, Fluorescein
Product Catalog No. FL-1061-5
Unit Size 5 mg
Lot No. ZK0523
Lot Expiration Date May 31, 2025

ASSAY	SPECIFICATION	LOT DATA
ACTIVITY	> 85%	PASS
TISSUE TEST	EQUIVALENT TO CONTROL	PASS

Quality Assurance Hayreck Kame

Date: 05/24/2023

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.