

# Certificate of Analysis



**Product Name** NanoLINK® Streptavidin Magnetic Beads (1.0 µm)  
**Product Catalog No.** M-1002-010  
**Unit Size** 1 ml  
**Lot No.** ZK0308  
**Lot Expiration Date** September 30, 2024

| ASSAY            | SPECIFICATION                                | LOT DATA |
|------------------|--|----------|
| BINDING CAPACITY | ≥ 12 nmol fluorescein-biotin per mg of beads | PASS     |
| SOLID CONTENT    | 10 mg/ml by A <sub>600</sub>                 | PASS     |

Quality Assurance

Date: 03/10/2023

*Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.*