

Certificate of Analysis



Product Name Streptavidin Agarose Ultra Performance™
Product Catalog No. N-1000-002
Unit Size 2 ml
Lot No. ZL0112
Lot Expiration Date August 31, 2024

ASSAY	SPECIFICATION	LOT DATA
BINDING CAPACITY	≥ 330 nmol fluorescein-biotin per ml resin	PASS

Quality Assurance

Hayrett Kams

Date:

1/16/2024

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.