

Certificate of Analysis



Product Name PNA, Rhodamine
Product Catalog No. RL-1072-5
Unit Size 5 mg
Lot No. ZK1024
Lot Expiration Date October 31, 2025

ASSAY	SPECIFICATION	LOT DATA
ACTIVITY	> 85%	PASS
TISSUE TEST	EQUIVALENT TO CONTROL	PASS

Quality Assurance

Harriet Kauer

Date:

10/24/2023

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.