



CERTIFICATE OF ANALYSIS

Product **FLUORESCCEIN STREPTAVIDIN**
Catalog No. SA-5001
Amount 1 mg active conjugate
Lot No. ZG0210
Concentration 1.0 mg/ml active conjugate
Solution 10 mM HEPES, 0.15 M NaCl, pH 7.5, 0.08% sodium azide.
Free Fluorescein None
F/P (molar) 4.6
Maximum Excitation @ 495-500 nm
Maximum Emission @ 514-521 nm
Active Conjugate 100%
Storage Conditions Refrigerate in the dark.

Other Remarks:

For diluting this product, we recommend a HEPES-, bicarbonate- or phosphate-buffered saline solution, approximately pH 8.2. Avoid using RPMI 1640 or other biotin-containing solutions as diluents. Serum also can contain biotin and should not be added to diluents.

The recommended range of concentrations for staining of cells or tissues is 5-30 µg/ml.