

# Certificate of Analysis



**Product Name** Streptavidin, Peroxidase, ELISA Grade  
**Product Catalog No.** SA-5014-1  
**Unit Size** 1 mg  
**Lot No.** ZK0622  
**Lot Expiration Date** June 2025

ASSAY	SPECIFICATION	LOT DATA
ELISA	EQUIVALENT TO CONTROL	PASS

Quality Assurance Harpreet Kaur

Date: 06/27/2023

Vector Laboratories certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.